



fluorescent foil

Fluorescent foil for the detection of xray radiation

Typically used for adjustment or monitoring of the primary beam

Wavelength (anode material): Cu, Mo, Co (Fe, Ag)

One side coated with fluorescent pigment the other side with sticky tape (Thickness 0.27mm including carrier foil, 0.12mm without carrier foil)

Size: A4 (others on request, max. 45x66 cm*cm)

